



# DBRS Medical Systems, Inc.

(A C.T. Consulting Group)

07204

13820 Benson Ave • Chino, CA 91710  
(909) 364-9547 • FAX (909) 548-4633

DATE: 5<sup>th</sup> Sep 2018

BILL TO:	ACCT #	SERVICE TYPE	DESCRIPTION
ATTN: Greenlight Imaging		Billed	Scheduled PM
NAME: 8348 Rosemead BLVD.	P.O. #	PHONE NUMBER	
STREET:		(562) 222-1321	
CITY: PICO RIVERA CA. 90660	CONTRACT I.D.	HOSPITAL DEPT.	ROOM
STATE:	Greenlight #4	Radiology	Mobile CT
ZIP:	SERVICE REQUESTED BY:		
	Greenlight Imaging		
MFG/MODEL	DESCRIPTION	READING	EXAMS
General Electric	Lightspeed Cxi	32613871	1839
		MAS	SCAN SEC.
			SLICE 146180
			GANTRY ROTATION 4040

**PROBLEM:** Scheduled PM. -

Tube warm up  
Test Scan  
Remove Replace DAS filters deteriorated

clean slip ring covers  
Repair control panel  
Replace Control Panel  
Remove table covers & cradle

Fix Loose light window  
Case lights good  
Needs new phantom

**CAUSE:** VAC DAS FANS  
VAC OBL FAN  
VAC SEC FAN  
VAC tube Hubs filters

Clean & Lub table  
VAC table base  
SBC screws caps r broken

Collimator Cal  
2 slope Cal  
Feed shift  
Test QA

**WORK PERFORMED:**

Open Col Window  
Clean Detector  
VAC entire gantry internally  
VAC entire gantry externally  
VAC all fan/ components in Gantry  
Clean Slip Rings & still stuck 2 little (be-wash sticks)  
Clean Slip Rings

VAC console  
Install console filter  
DARK DRIVE Hours are A-30,284  
HP Drives A-259  
B-972  
C-143

TUBE WARM UP  
Test Scan  
QA  
DAS Gain Cal  
to fix mylar  
window error  
Inspect mylar  
window

DATE:	EMPLOYEE NUMBER	TRAVEL HOURS	HOURS WORKED			MILES	RATE
			STD.	O.T.	O.T. PREM.		
5 <sup>th</sup> Sep 18	9989	0	5	1.5		0	
	0929	0	2				
<b>TOTALS</b>		0	7	1.5		0	Bill @ Warehouse Rates

QTY.	PART NUMBER	DESCRIPTION	LIST
1	2252889	SDAS FILTERS CENTER	\$45.00
1	2252889-2	LIR SDAS FILTERS	\$35.00
1	5128180	GRE FRONT COVER FILTER	\$105.00
1	2112363	Gantry Control Switch Panel SN#0211	\$550.00
1	2117929	Gantry Control Rest Cap	\$25.00
1	2282912	Gantry Control Panel LED	\$22.00

**VERIFIED**

DATE COMPLETED: 5<sup>th</sup> SEP 2018  
DBRS SIGNATURE: [Signature]  
CUSTOMER SIGNATURE: [Signature]

EQUIPMENT (OUT) DATE 9/5/18 OWNER (DBRS/CUSTOMER) DBRS COST \_\_\_\_\_ EXCHG X PURCH \_\_\_\_\_  
PART # 2112363 S/N 0211 SHIP VIA At office LOCATION \_\_\_\_\_ SIGN [Signature]  
SHIPPED TO Greenlight CTY RESTOCK \_\_\_\_\_ REPAIR \_\_\_\_\_ INVEN \_\_\_\_\_  
EQUIPMENT (IN) DATE 9/6/18 S/N 0245 RESTOCK \_\_\_\_\_ REPAIR \_\_\_\_\_ INVEN \_\_\_\_\_  
PART # 2112363 S/N 0245 RESTOCK \_\_\_\_\_ REPAIR \_\_\_\_\_ INVEN \_\_\_\_\_  
RECEIVED FROM Greenlight CTY SHIP VIA AT office LOCATION 60 VERIFIED [Signature]

**VERIFIED**

DESC. Gantry control Switch Panel