

Schedule E Planned Maintenance

Site: <u>Brockton</u>	Month: <u>10/2021</u>
System ID#: _____	Dispatch/Job #: _____
Service Engineer: <u>MK</u>	

PM Check	Test Purpose	Pass	Fail	Comments
Image Quality				
<u>Alignment Light Check</u>	IQ	/		
<u>LVShim</u>	IQ	/		
<u>Grafidy3</u>	IQ	/		
<u>Coherent Noise</u>	IQ	/		
<u>Signal to Noise Check</u>	IQ	/		
<u>Spike Noise Check</u>	IQ	/		
<u>PM Check (Service Contract Customers Only)</u>	IQ			
Magnet Room				
<u>Patient Blower & Filter</u>	Check	/		<u>Clean</u>
RF/System Cabinet				
<u>Fans & Filter for SRFD</u>	Clean	/		
<u>Check RF Output Power</u>	Check	/		
Patient Handling				
<u>Patient Table Checks</u>	Safety	/		
Gradients				
<u>Fans & Filter</u>	Clean	/		
<u>Gradient Cable/PDU Connections and Support</u>	Replace	/		
<u>Check Fluid Levels and Valve of Heat Exchanger</u>		/		
Twin Accessory Cabinet				
<u>Filter Replacement</u>	Replace			<u>N/A</u>
Computer				
<u>Storelog</u>	Check	/		
Magnet				
<u>Test GE Magnet Rundown Unit (MRU)</u>	Safety	/		

Schedule E Signa Planned Maintenance.

Date Completed: 10/10/2021