



Scheduled Planned Maintenance

Site: Greenlight Imaging MRI **Month:** 03/15/2024

System ID: 12345 **Service Engineer:** Alex Gonzalez

PM Check	Test	Pass	Fail	Comments
----------	------	------	------	----------

IMAGE QUALITY

Alignment Light Check	IQ	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
DC Offsets	IQ	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
DQA Calibration	IQ	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Head and Body SNR	IQ	<input checked="" type="checkbox"/>	<input type="checkbox"/>	Adjusted head and body
LV Shim	IQ	<input checked="" type="checkbox"/>	<input type="checkbox"/>	In spec
SPT Calibration	IQ	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

MAGNET ROOM & PATIENT HANDLING

Patient Cradle	Check	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Patient Table	Check	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Emergency Stop Button	Check	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Test MRU	Check	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Inspect MRU	Check	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

RF / SYSTEM & GRADIENT CABINET

Filters in RFS Cabinet	Clean	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Fans & Filter for SRFD	Clean	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Check RF Output Power	Clean	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Gradient Filters	Clean	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

COMPUTER

Storelog	Check	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Clean Dust , Set Time	Clean	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Pheripheral SCSI Tower	Clean	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Perform SaveINFO	N/A	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

Date Completed: 03/15/2024 **Engineer:** A.G